

J.I. 0D8A1.AX
(3610, 3611)

April 4, 1998

STRAINGAGE TRANSDUCER, L-CELL
FOR
HAZARDOUS AREAS

from

KISTLER-MORSE CORPORATION
19021-120TH AVE. N.E.
BOTHELL, WA 98011-9511**I. INTRODUCTION****1.1 Standards**

KISTLER-MORSE CORPORATION requested Approval of the apparatus listed in Section 1.2 to be in compliance with the applicable requirements of the following standards:

<u>Title</u>	<u>No.</u>	<u>Issue Date</u>
Intrinsically Safe Apparatus for Use in Class I, II and III, Division 1 Hazardous (Classified) Locations	FMRC 3610	October 1988
<i>Note: 1.5 factor applied to voltage and current rather than energy.</i>		
Electrical Equipment for Use in Class I, Division 2 Class II, Division 2 and Class III, Division 1 and 2 Hazardous Locations	FMRC 3611	April 1986
Electrical Equipment for Use in Hazardous (Classified) Locations General Requirements	FMRC 3600	December 1999
Electrical and Electronic Test, Measuring, and Process Control Equipment	FMRC 3810	March 1989
Enclosures for Electrical Equipment (1000 V max)	NEMA Pub. 250	1991